Product Instructions

1.1 Model

ResAssist for Web, ResAssist for App (Android), ResAssist for App (IOS)

1.2 Intended Use

ResAssist is intended to receive patients data from Resvent compatible devices (Positive Airway Pressure System) via SD cards/Wifi/GSM and process them into dat a reports. When doctors make healthcare decisions, the data can be used as one of the references. ResAssist can be used in Web and application. The software in Web can r emotely change the parameters of Resvent compatible devices(e.g. ventilation mode, t reatment pressure, humidity, etc.) to treat obstructive sleep apnoea(OSA) under the a ccounts of doctors.

1.3 Contraindications

Refer to the user guide of the attached Resvent compatible devices (Positive Air way Pressure System) for contraindications or adverse effects.

1.4 User qualifications

The person operating the device by the instruction in the user’s manual is referred as the “user”. In contrast, a “patient” is the person receiving the therapy. Always perform all the operating steps in accordance with the user’s manual.

Clinicians can receive specified professional training from the manufacturer-authorized distributor regarding proper device operation.

1.5 Important

Read this entire guide before using the device.